

PCE Americas Inc. 711 Commerce Way Suite 8 Jupiter FL-33458 USA From outside US: +1 Tel: (561) 320-9162 Fax: (561) 320-9176 info@pceamericas.com PCE Instruments UK Ltd. Units 12/13 Southpoint Business Park Ensign way Hampshire / Southampton United Kingdom, SO31 4RF From outside UK: +44 Tel: (0) 2380 98703 0 Fax: (0) 2380 98703 9 info@pceinstruments.com

www.pce-instruments.com/english www.pce-instruments.com

PCE PCE-PQC 12EU Particle Counter

$\begin{array}{c} \mbox{Particle counter PCE-PQC 1xEU series incl. calibration certificate} \\ \mbox{Measurement of particle sizes up to 25 μm / Up to 3 parallel measuring channels / Internal memory / Reporting according to ISO 14644-1, EU GMP Annex I, FS 209E / Extrapolation of the mass concentration in μg / m³ / Ethernet, USB or (optional) Wifi connection \\ \end{array}$

The particle counters of the PCE-PQC 1xEU series measure the concentration of particles such as dust, soot, pollens and many other aerosols in the air. To accurately determine the degree of pollution of the air, the particle counters were developed. Pollution is mainly generated by combustion, material processing, manufacturing, power generation, vehicle engine emissions and the construction industry. With the help of the PCE-PQC 1xEU series particle counters the exact amount of dirt particles in the air can be measured.

Of greater importance is the degree of pollution of the air with pathogenic particles such as soot, which are released by the industry and especially by diesel vehicles without special filtering in the air. Such dispersion particles are responsible inter alia for reduced visibility, the inhalation of toxic substances and thus reduced labor productivity. It has also been known for some time that the particles also make a not inconsiderable contribution to many medical diseases such as asthma, bronchitis, skin and lung diseases. The particle counters are designed for easy and quick use. The particle counters of the PCE-PQC 1xEU series operate in different modes (real-time, cumulative, differential, mass concentration, ...) and display the results on the display,

- Internal memory
- Particle sizes up to 25 µm
- Color display
- Handy
- 3 measuring channels
- ISO 14644-1, EU GMP Annex I, FS 209E
- Mass concentration
- Ethernet, USB
- Incl. calibration certificate traceable to NIST ISO 21501-4 and JIS B9921

SPECIFICATION

Measuring range Measuring channel sizes Counting efficiency 0.3 ... 25 μm Factory calibrated at 0.3, 0.5, 5.0 μm 50% at 0.5 μm 100% at > 0.75 μm according to JIS Flow Random loss Battery Light source Zero count

Counting modes

Alarms Calibration Display Printer Aspiration

Air outlet Battery pack Charging time Reports

Configuration Standards Dimensions Weight Storage

Sample locations Samples duration Power supply Operating conditions

Storage conditions

Interface Optional interfaces Number of measuring channels 2.83 l / min (0.1 ft3 / min) 5% at 4,000,000 particles / ft3 10 h Long-life laser diode < 1 count / 5 min (< 2 particles / ft³) According to ISO 21501-4 and JIS Automatic, manual, real-time, cumulative / Differential, mass concentration 1 ... 9999999 counts, adjustable Traceable to NIST 4.3" WQVGA color touch display, 480x272 px External thermal printer Internal pump with automatic Flow control Internal HEPA filter Replaceable Li-Ion battery About 4 hours ISO 14644-1 EU GMP Annex 1 FS 209E Memory for 50 custom configurations ISO 21501-4 and JIS B9921 25.4 x 12.9 x 11.4 cm 1.0 kg 45000 data sets (ring memory) consisting of Particle count, temp and humidity, locations and times Up to 1000 locations can be stored 1 s ... 99 h adjustable 110 ... 240V AC 50/60 Hz 5 ... 40°C / 41 ... 104°F Up to 95% RH not condensing 0 ... 50°C / 32 ... 122°F Up to 98% RH not condensing

USB

Ethernet, WiFi 802.11 b / g, RS485 or RS232 3

DELIVERY SCOPE

- 1 x Particle counter PCE-PQC 12EU
- 1 x Isokinetic probe
- 1 x Cleaning filter
- 1 x Li-Ion rechargeable battery
- 1 x Software for data download
- 1 x USB cable
- 1 x Power supply / battery charger
- 1 x User manual
- 1 x Calibration certificate according to NIST