



## C 7000 basic equipment set 2

/// Data Sheet

The C 7000 is the first IKA® calorimeter with a completely dry system for measuring the gross calorific value of solid and liquid samples. The temperature is measured directly in the decomposition system. This results in measurement times in the range of 3 to 7 minutes (depending on the sample). The system can manage up to 8 different decomposition vessels using a code ring scheme.

- High sample frequency
- Precise and reproducible determination of gross calorific values according to ISO 1928
- Reduction of routine work through automatic application flow
- Automatic decomposition vessel identification



designed for scientists

- Interface connections for scale, printer and PC
  - User-friendly software C 5040 CalWin for controlling the calorimeter and administration of measuring data
  - Special halogen resistant vessel for quantitative decomposition of halogens and sulfur
  - The decomposition vessel can be changed over to use combustible crucibles C 14
- C 7000 basic equipment set 2 consisting of:
- C 7000 Measurement cell
  - C 7012 Decomposition vessel, halogen resistant
  - C 7002 Cooling system
  - C 48 Oxygen station



designed for scientists

## Technical Data

Measuring range max. [J]	30000
Measuring mode double dry (ISO 1928)	yes
Measuring time double dry approx. [min]	3
Working temperature max. [°C]	30
Temperature measurement resolution [K]	0.0001
Cooling medium temperature [°C]	12 - 30
Cooling medium permissible operating pressure [bar]	9
Cooling medium	tap water
Type of cooling	flow
Flow rate [l/h]	120 - 180
Oxygen operating pressure max. [bar]	40
Interface scale	RS232
Interface printer	Centronix
Interface PC	RS232
Decomposition detection	yes
Decomposition vessel C 7012	yes
Decomposition vessel, halogen resistant	yes
Dimensions (W x H x D) [mm]	800 x 395 x 490
Weight [kg]	50.207
Permissible ambient temperature [°C]	18 - 30
Permissible relative humidity [%]	80
RS 232 interface	yes
Voltage [V]	220 - 240
Frequency [Hz]	50/60
Power input [W]	260