

User Manual

PCE-GM 75 / PCE-GM 80 Gloss Meter



User manuals in various languages (français, taliano, español, português, nederlands, türk, polski, русский, 中文) can be found by using our

product search on: www.pce-instruments.com

Last change: 28 February 2023 v1.0

© PCE Instruments



Contents

1	Safety notes	1
2	Instrument Characteristics	2
3	Specifications	2
4	Delivery contents	2
5	Display descriptions	2
6	Operation instructions	3
7	Menu instructions	3
8	Modifying the calibration value	3
9	Setting the time	4
10	Important notice	4
11	Contact	4
12	Disposal	4

1 Safety notes

Please read this manual carefully and completely before you use the device for the first time. The device may only be used by qualified personnel and repaired by PCE Instruments personnel. Damage or injuries caused by non-observance of the manual are excluded from our liability and not covered by our warranty.

• The device must only be used as described in this instruction manual. If used otherwise, this can cause dangerous situations for the user and damage to the meter.

PCE

- The instrument may only be used if the environmental conditions (temperature, relative humidity, ...) are within the ranges stated in the technical specifications. Do not expose the device to extreme temperatures, direct sunlight, extreme humidity or moisture.
- Do not expose the device to shocks or strong vibrations.
- The case should only be opened by qualified PCE Instruments personnel.
- Never use the instrument when your hands are wet.
- You must not make any technical changes to the device.
- The appliance should only be cleaned with a damp cloth. Use only pH-neutral cleaner, no abrasives or solvents.
- The device must only be used with accessories from PCE Instruments or equivalent.
- Before each use, inspect the case for visible damage. If any damage is visible, do not use the device.
- Do not use the instrument in explosive atmospheres.
- The measurement range as stated in the specifications must not be exceeded under any circumstances.
- Non-observance of the safety notes can cause damage to the device and injuries to the user.

We do not assume liability for printing errors or any other mistakes in this manual.

We expressly point to our general guarantee terms which can be found in our general terms of business.



2 Instrument Characteristics

- Fine, light, stylish appearance and easy to carry
- Completely intelligenised design, single key operation, convenient to use
- Automatically adjustable
- Capable of measuring from multiple different angles
- Long lifespan lamp
- High measurement precision
- Backlit LCD digital display with accurate battery life and time display
- Auditory feedback during operation
- Automatic shutdown
- USB interface for PC connection
- Stores up to 1245 measurements

3 Specifications

Measurement range	0 99.9 GU		
	100 1000 GU		
Resolution	0.1 GU		
	1 GU		
Repeatability	< 0.45 GU		
	0.5%		
Accuracy	< 1 GU		
	1%		
Measurement geometry	60°		
Measuring surface	Ø4.5 mm		
Data storage	1,245 measured values		
Interface	Micro USB		
Power supply	2 x 1.5V AAA batteries		
Environmental conditions	0 40°C / 32 313°F, < 85% rh non-condensing		
Dimensions	115 x 65 x 25 mm / 4.5 x 2.6 x 1 in		
Sensor cable length (PCE-GM 80 only)	approx. 50 cm		
Weight	100 g / < 1 lb		
Standards	GB9754-88, GB9966.5, ISO2813, JJG696-2002		

4 Delivery contents

- 1 x gloss meter PCE-GM 75 or PCE-GM 80
- 1 x external sensor (PCE-GM 80 only)
- 1 x calibration standard
- 1 x micro USB cable
- 1 x cleaning cloth
- 2 x 1.5 V AAA battery
- 1 x user manual

5 Display descriptions

- "GU" Gloss Unit
- "calibrate" Runs the calibration process
- "clear save" Clears the saved measurements
- "Change CAL" Modifies the calibration value
- "Set time" Modifies the time display



6 Operation instructions

Press the button to start the device. The display will now show "Calibrate?" Place the device, with the measurement window into the calibration slot, in the case of the device and press the button again to confirm the calibration process. The device now calibrates automatically. Once calibrated, the display shows "calibrate ok" and then "ready". The device is now in the working condition and is ready for use.

(If you choose the skip the calibration process, simply wait ~20 sec. until the display shows "ready".)

Now place the device with the measurement window onto the surface you wish to measure and press the button to begin. With a "beep" the device displays the gloss value of the measured surface.

The device will shut down automatically after ~50 sec. of not being used.

7 Menu instructions

When the device is ready for use, press and hold the button until the display changes and "SET/MENU" is highlighted. Release the button to enter the next screen. The display change looks as follows:



The cursor now cycles (top to bottom) the options of the next menu. Press the button when the desired option is highlighted in order to select:



8 Modifying the calibration value

If you wish to replace the standard plate (black quartz crystal plate), you must then change the calibration value. In order to do this, you must press the button when "change CAL" is highlighted inside the "SET/MENU" screen (refer to picture above). The display now shows "change CAL?" Press the button again to confirm and begin the process (if no new value is needed, do not press the button, after ~20 sec. the device will return to the screen displaying "ready"). The device now begins to count *up* from 80. Press the button at the desired value to confirm it. The device will now return to the working condition and display "ready".





If you wish to change the time displayed on the device, press the button when "set time" is highlighted inside the "SET/MENU" screen (refer to picture in point 7 Menu Instructions). The highlighted area (beginning with the year) now begins to count up, press the button again to confirm the desired value. After confirmation, the next value will be highlighted and begin to count up (month). Repeat the process until you have the desired date and time set.

10 Important notice

- Avoid direct sunlight; otherwise it can affect the measurement accuracy.
- The standard plate should be kept clean. Avoid touching it with your fingers. If there is dust on the plate, it can affect the measurement accuracy. Wipe it with the provided cloth or absolute alcohol.
- When the battery level is less than 20%, the batteries need to be replaced. Open the
 battery cover and replace the old batteries with new ones. Please note the polarity of
 the contact wires.
- The batteries should be replaced when depleted, otherwise long-term storage may cause liquid to leak out and damage the device. If the device will not be used for a longer period of time, the batteries should be taken out and kept in a safe place.

11 Contact

If you have any questions, suggestions or technical problems, please do not hesitate to contact us. You will find the relevant contact information at the end of this user manual.

12 Disposal

For the disposal of batteries in the EU, the 2006/66/EC directive of the European Parliament applies. Due to the contained pollutants, batteries must not be disposed of as household waste. They must be given to collection points designed for that purpose.

In order to comply with the EU directive 2012/19/EU we take our devices back. We either re-use them or give them to a recycling company which disposes of the devices in line with law.

For countries outside the EU, batteries and devices should be disposed of in accordance with your local waste regulations.

If you have any questions, please contact PCE Instruments.





PCE Instruments contact information

Germany

PCE Deutschland GmbH Im Langel 26 D-59872 Meschede Deutschland Tel.: +49 (0) 2903 976 99 0 Fax: +49 (0) 2903 976 99 29 info@pce-instruments.com www.pce-instruments.com/deutsch

United Kingdom

PCE Instruments UK Ltd Unit 11 Southpoint Business Park Ensign Way, Southampton Hampshire United Kingdom, SO31 4RF Tel: +44 (0) 2380 98703 0 Fax: +44 (0) 2380 98703 9 info@pce-instruments.co.uk www.pce-instruments.com/english

The Netherlands

PCE Brookhuis B.V. Institutenweg 15 7521 PH Enschede Nederland Telefoon: +31 (0)53 737 01 92 info@pcebenelux.nl www.pce-instruments.com/dutch

France

PCE Instruments France EURL 23, rue de Strasbourg 67250 Soultz-Sous-Forets France Téléphone: +33 (0) 972 3537 17 Numéro de fax: +33 (0) 972 3537 18 info@pce-france.fr www.pce-instruments.com/french

Italy

PCE Italia s.r.l. Via Pesciatina 878 / B-Interno 6 55010 Loc. Gragnano Capannori (Lucca) Italia Telefono: +39 0583 975 114 Fax: +39 0583 974 824 info@pce-italia.it www.pce-instruments.com/italiano

United States of America

PCE Americas Inc. 1201 Jupiter Park Drive, Suite 8 Jupiter / Palm Beach 33458 FL USA Tel: +1 (561) 320-9162 Fax: +1 (561) 320-9176 info@pce-americas.com www.pce-instruments.com/us

Spain

PCE Ibérica S.L. Calle Mayor, 53 02500 Tobarra (Albacete) España Tel. : +34 967 543 548 Fax: +34 967 543 542 info@pce-iberica.es www.pce-instruments.com/espanol

PCE

Turkey

PCE Teknik Cihazları Ltd.Şti. Halkalı Merkez Mah. Pehlivan Sok. No.6/C 34303 Küçükçekmece - İstanbul Türkiye Tel: 0212 471 11 47 Faks: 0212 705 53 93 info@pce-cihazlari.com.tr www.pce-instruments.com/turkish

Denmark

PCE Instruments Denmark ApS Birk Centerpark 40 7400 Herning Denmark Tel.: +45 70 30 53 08 kontakt@pce-instruments.com ww.pce-instruments.com/dansk